

Bradford-Assay**Standards:**

		BSA (10 mg/ml)	H ₂ O
2.5	mg/ml	25.0 μ l	75.0 μ l
1.25	mg/ml	12.5 μ l	87.5 μ l
1.0	mg/ml	10.0 μ l	90.0 μ l
0.75	mg/ml	7.5 μ l	92.5 μ l
0.5	mg/ml	5.0 μ l	95.5 μ l
0.25	mg/ml	2.5 μ l	97.5 μ l
0	mg/ml	0 μ l	100.0 μ l

10 μ l standard/sample
+ 1ml Bradford reagent
Incubate at RT 5-10 min
Measure at 595 nm

Bradford-reagent:

100 mg Coomassie Blue **G250**
+ 50 ml **95% Ethanol** p.A.
+ 100 ml **85% H₃PO₄**
ad 1000 ml H₂O

Filter Bradford-solution and store at 4°C in the dark.

10 mg/ml BSA Standard (Sigma) in H₂O